

**CERTIFICATE
OF COMPLIANCE
USP BIOLOGICAL TESTS
CLASSIFICATION VI**

Material tested: Elastosil R401/40 Control No: 33932HO
Manufacturer: WACKER Chemie GmbH, Werk Burghausen
Johannes-Hess-Str. 24, D-84489 Burghausen
Studies performed:

ACUTE SYSTEMIC TOXICITY (USP)

The saline, alcohol in saline, polyethylene glycol 400 and vegetable oil extracts of the test item injected into mice did not produce a significantly greater systemic reaction than the blank extractant.

INTRACUTANEOUS TOXICITY (USP)

The saline, alcohol in saline, polyethylene glycol 400 and vegetable oil extracts of the test item injected intracutaneously in rabbits did not produce a significantly greater tissue reaction than the blank extractant.

IMPLANTATION TEST (USP)

The macroscopic reaction of the test item implanted 120 hours was not significant as compared to the USP negative control plastic.

The sample of the test item extracted at a ratio of 60 cm²/20ml and at a temperature of 50 ± 1°C for 72 hours met the requirements of a USP Class VI Plastic.

BIOSERVICE Scientific Laboratories GmbH Munich

Behringstraße 6

D-82152 Planegg

Albrecht Poth

Dr. Albrecht Poth
Manager Biological Safety Testing



Date: July 10, 1998